



TEST REPORT TUCHENG

Date of Issue: Dec.07,2020 Date of Receipt and Test Start: Nov.20,2020

Report No.: TFF9K489 Quantity: IPC Page Order/Pages: (P1/3) Ref. No.: NIL

Report Title: FOXLUX CO., LTD.(U3454) Item: Mask

Address: No. 38, Ln. 838, Zhuangjing Rd., Zhongli Dist., Taoyuan City 320, Taiwan (R.O.C.)

Test Items		Test Results	Test Methods
*Air Exchange Pressure (mmH2O/cm ³)	1	4.4	CNS 14774 T5017-2018 9.4
	2	4.3	CNS 14777 T4039-2003
	3	4.1	
	4	3.5	
	5	4.0	


This test report is for reference only and should not be used as a formal document.

Note: Air Exchange Pressure takes 5 masks for testing.

Note: Sample description is given by the client: Foxlux Medical Mask (Non-sterile)

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: "*" indicates that the tests are accredited by the Ministry of Health and Welfare in Taiwan, and are conducted according to the accredited test methods.



- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of Jui-hung kao
Taiwan Textile Research Institute
Director,
Department of Testing and Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

Date of Issue: Dec.07,2020 Date of Receipt and Test Start: Nov.20,2020

Report No.: TFF9K489 Quantity: IPC Page Order/Pages: (P2/3) Ref. No.: NIL

Report Title: FOX LUX CO., LTD.(U3454) Item: Mask

Address: No. 38, Ln. 838, Zhuangjing Rd., Zhongli Dist., Taoyuan City 320, Taiwan (R.O.C.)

Test Items		Test Results	Test Methods
*Bacterial Filtration	1	99.0	CNS 14774-T5017-2018 9.2
Efficiency (BFE)(%)	2	98.8	CNS 14775 T4037-2003
Staphylococcus aureus	3	99.0	This test report is for reference only and should not be used as a formal document.
ATCC 6538	4	99.0	
	5	99.0	

Note: Control average: 2389 CFU.

Note: Mean particle size: 3.1 μm.

Note: Testing side: outside of specimen.


Note: Testing area: 39.5 cm².

Note: Flow rate : 28.3 L/min.

Note: Sample description is given by the client: Foxlux Medical Mask (Non-sterile)

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: "*" indicates that the tests are accredited by the Ministry of Health and Welfare in Taiwan, and are conducted according to the accredited test methods.



- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
- 2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
- 3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of Jui-hung kao
Taiwan Textile Research Institute
Director,
Department of Testing and Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839

TEST REPORT TUCHENG

Date of Issue: Dec.07,2020 Date of Receipt and Test Start: Nov.20,2020

Report No.: TFF9K489 Quantity: 1PC Page Order/Pages: (P3/3) Ref. No.: NIL

Report Title: FOXLUX CO., LTD.(U3454) Item: Mask

Address: No. 38, Ln. 838, Zhuangjing Rd., Zhongli Dist., Taoyuan City 320, Taiwan (R.O.C.)

*This test report is for reference only
and should not be used as a formal
document.*



- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung kao
Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839